

Indira Gandhi Medical College & Research Institute

(Government of Puducherry Institution)

Vazhudavur Road, Kathirkamam, Puducherry - 605 009 Ph :0413-2277545, Fax:0413-2277289

Email: puducherrymedicalcollege@gmail.com, Web:www.igmcri.com

No.515/IGMC&RI/Estt-Pur/A5/Microbiology(part1)/2021-22 /2411

Date:

- 1 OCT 2021.

Calling for Quotation

Sub: IGMC&RI -Purchase- UV box(UV-Thermal Treatment)kit item - Reg.

The following item is proposed to be purchased of this Institution. It is therefore requested to quote the lowest possible rate for the supply of the following items. The **Hard Copy** of the quotation shall be submitted to the **Director, IGMC&RI, Puducherry. on or before 13.09.2021.**

Sl. No	Name of the Equipments	Qty
1	UV box(UV-Thermal Treatment)kit (Technical Specification Enclosed)	1No.

Terms and conditions:

1. **Prices should be inclusive of all taxes.**
2. The supplier should be responsible until the items arrived to this Institution in good condition
3. The rates once quoted and approved will be final and no revision in rates will be allowed
4. Supply should be made within 15 days from the date of receipt of our supply order
5. Delivery should be made on FOR, Puducherry, door delivery basis.
6. The quotation should be based strictly on the specification.
7. The material should be supplied strictly in accordance with the specifications enclosed/ mentioned. The items which are not conforming to the specifications should be taken back at their cost.
8. The quotation shall contain the GST Number, Contact Phone Number & E Mail I.D.
9. The quotation cover should be properly sealed with postal address.

// BY ORDER//

Yours faithfully,



ADMINISTRATIVE OFFICER

To:

1. ~~The Asst. Programmer, IGMC&RI, Puducherry- with instruction to upload in the official Web-Site of this Institute and for compliance report.~~

Copy to:

1. The HOD.,(Microbiology), IGMC&RI, Puducherry.

UV Sterilizer Chamber Technical Specification

1. Outer MS with white powder coated
2. Five tray white powder coated
3. OSRAM/Philips UVC TUBE 5No
4. Ozone (O3 Generator) for sterilization of shadow places
5. Six reflective surface for uniform exposure of UV rays
6. Front loading
7. Convenient pull out trays
8. Digital Timer
9. Door Safety Switch
10. UVC light cut of sensor
11. Odour exhaust
12. Auto UVC light expiry sensor
13. One Years Warranty



Dr. S. SRINIVASAN, MD.
PRINCIPAL INVESTIGATOR
Virus Research & Diagnostic Laboratory,
Indira Gandhi Medical College &
Research Institute, Puducherry - 605 009.